بسم الله الرحمن الرحيم

**Ministry of Finance**
Ameenee Magu, Male', Republic of Maldives
Telephone: (960) 334 9266, Fax: (960) 332 0706

މިނިސްޓްރީ އޮފް ފިނޭންސް
އަމީނީމަގު ، މާލެ، ދިވެހިރާއްޖެ

# CLARIFICATION 1

ކްލެރިފިކޭޝަން 1

| | | |
|---|---|---|
| ނަންބަރު<br>**No:** | **TES/2020/G-008** | |
| ޕްރޮޖެކްޓް<br>Project: | **Supply and Delivery of Trauma Implants** | |
| ތާރީޚް<br>Issued Date | **13th August 2020** | |
| ސަފްހާގެ އަދަދު<br>No. of Pages: -01 | ބީ.އޯ.ކިއު<br>BoQ: -00 | ޑްރޯއިން<br>Drawings: -00 |

Please include this clarification when submitting the bid .ބީޑް ހުށަހަޅުއްވާއިރު މި ކްލެރިފިކޭޝަން ޑޮކިއުމަންޓާއި އެކު ހުށަހަޅުއްވަން ޖެހިވަޑައިގަންނަވާނެއެވެ

- **Find the below Clarifications for the Queries raised**

| **No.** | **Queries** | **Clarification** |
|---|---|---|
| 1- | **Product material:** Normally orthopedic implants are made with two types of materials, medical-grade stainless steel or Titanium, these two are very different in terms of material quality and there is a big difference between the prices of these two. Therefore, these two products cannot be considered the same in an evaluation of a tender like this one. Therefore, it is very important to specify the material type of the implant specifically plates and screws, which goes to the human body. | Material specification are for Titanium products. |
| 2- | **Product specification:** The specification or product description given in this tender is specific to a particular brand/manufacture, therefore would like to know if the similar / equivalent from other manufactures would be acceptable. | Yes,<br>Similar / equivalent from other manufacturers can be accepted. Parties need to provide 2 Nos of Instruments sets (FOC) for the implants (brand) offered. |
| 3- | **Product certification/quality assurance certifications:** As highlighted there are two major medical related certifications used worldwide either CE (Europian standard) or FDA ( USA standard), either of these two will assure product quality, therefore it will be easier for a fair and comprehensive evaluation if you include product quality certification in your requirement. | Either CE or FDA certifications should be provided along with all the data sheets / brochures and pictures of the implants. |

| | |
|---|---|
| ނަން<br>**Name: Fathimath Rishfa Ahmed** | ސޮއި<br>**Signature:** |